



SAP / Monitoring



Bowie & Dick Helix for Hollow Loads

REF: 101.001.0100 Hollow A Helix with 1,5 mtr tube Unit: 100 indicators

REF: 101.002.0250 Hollow A Helix with 1,5 mtr tube Unit: 250 indicators

Helix devices and their indicators have proven that they detect failures during each cycle much earlier than other chemical indicators / integrators.

The EN 285 + Amendment A2: 2009 are specifying the Helix test (Hollow A) to be used as Bowie Dick Test when hospitals are sterilizing Hollow Loads.

Next to that this Bowie Dick Helix Test is an obligatory test for Class B Bench Top Steam sterilizers.

Characteristics:

- Conform to ISO 11140-1 Class 2 and ISO 11140-4 : Class 2
- Conform to EN 867 Part 5 Class B
- Clear transition of colour into black
- Easy to interpret
- Easy to document
- Non-Toxic / Lead Free
- Can be fully recycled

